
Midmark M11 - v1.0.5
Total Cycles: 1376

Name: RITTER2
Sterilizer ID: V2313228
08/08/2022 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.9 F 0.0
2:00 178.6 F 0.2
4:00 196.4 F 4.4
6:00 214.3 F 9.8
8:00 227.5 F 7.7
10:00 238.0 F 12.2
12:00 238.9 F 10.9
14:00 238.7 F 10.3
16:00 253.1 F 17.5
18:00 265.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.7 F 29.7
2:00 272.0 F 29.0
3:00 271.9 F 29.5
4:00 272.7 F 29.7

Min 271.5 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.8 F 0.0
5:00 215.7 F 0.0
10:00 223.6 F 0.0
15:00 223.1 F 0.0
20:00 221.9 F 0.0
25:00 220.8 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

08/08/2022 02:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____