
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000785
12:26 10/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:47
136.4 C 223kPa 4:47
136.4 C 224kPa 5:17
136.2 C 222kPa 5:47
136.4 C 225kPa 6:17
136.4 C 225kPa 6:47
136.3 C 223kPa 7:17
136.1 C 220kPa 7:47
136.4 C 222kPa 8:17
136.4 C 222kPa 8:17
VENTING 8:17
AIR DRYING 9:12
CYCLE COMPLETE 9:16

Digital Signature #
DFF47759BF26BDB4
