

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4733

Name: RITTER 1  
Sterilizer ID: V2178198  
07/16/2024 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.2 F 0.0  
2:00 129.8 F 1.2  
4:00 148.3 F 4.2  
6:00 179.4 F 7.5  
8:00 192.1 F 11.0  
10:00 207.5 F 3.9  
12:00 223.4 F 6.0  
14:00 231.9 F 9.9  
16:00 235.7 F 10.1  
18:00 240.3 F 10.7  
20:00 235.6 F 8.3  
22:00 248.4 F 14.4  
24:00 259.2 F 20.3  
26:00 269.1 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.6 F 29.1  
2:00 272.3 F 28.7  
3:00 271.2 F 28.2  
4:00 272.1 F 29.0

Min 270.9 F 28.1  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
5:00 215.8 F 0.0  
10:00 185.0 F 0.0  
15:00 183.1 F 0.0  
20:00 180.2 F 0.0  
25:00 178.1 F 0.0  
30:00 179.4 F 0.0

Drying Complete

FILLING: 0:49  
HEATING: 27:13  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:05:58  
CYCLE COMPLETE

07/16/2024 11:11 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_