
Midmark M11 - v1.0.0
Total Cycles: 5311

Name: RITTER 1
Sterilizer ID: V2178198
02/03/2025 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.1 F 0.0
2:00 108.0 F 1.4
4:00 144.7 F 4.9
6:00 175.3 F 9.4
8:00 192.8 F 5.2
10:00 220.7 F 6.4
12:00 234.9 F 11.0
14:00 229.1 F 6.6
16:00 243.6 F 12.7
18:00 240.5 F 10.6
20:00 254.2 F 17.4
22:00 266.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.5
2:00 272.4 F 28.9
3:00 271.9 F 28.1
4:00 272.8 F 28.9

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 202.1 F 0.0
10:00 192.4 F 0.0
15:00 200.2 F 0.0
20:00 211.2 F 0.0
25:00 212.7 F 0.0
30:00 219.1 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 23:33
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:27
CYCLE COMPLETE

02/03/2025 05:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____