
Midmark M11 - v1.0.5
Total Cycles: 827

Name: RITTER2
Sterilizer ID: V2313228
02/23/2022 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.2 F 0.0
2:00 184.0 F 4.2
4:00 204.2 F 9.9
6:00 221.8 F 6.4
8:00 234.9 F 11.7
10:00 236.6 F 10.0
12:00 236.6 F 8.9
14:00 251.5 F 16.4
16:00 263.7 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.5 F 29.7
2:00 272.0 F 29.0
3:00 271.9 F 29.4
4:00 272.5 F 29.4

Min 271.3 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 220.8 F 0.0
10:00 221.9 F 0.0
15:00 225.4 F 0.0
20:00 223.5 F 0.0
25:00 223.7 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 17:54
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:15
CYCLE COMPLETE

02/23/2022 01:18 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____