

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2696

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2022 09:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.8 F 0.0  
2:00 80.5 F 0.0  
4:00 91.0 F 0.0  
6:00 116.3 F 0.7  
8:00 146.0 F 2.4  
10:00 173.7 F 5.2  
12:00 195.1 F 8.7  
14:00 210.5 F 12.4  
16:00 219.1 F 5.1  
18:00 230.0 F 9.2  
20:00 239.6 F 13.6  
22:00 238.7 F 10.2  
24:00 235.5 F 8.1  
26:00 248.3 F 14.2  
28:00 259.4 F 20.4  
30:00 270.1 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.8  
2:00 271.5 F 28.4  
3:00 272.4 F 28.7  
4:00 271.5 F 28.3  
  
Min 271.4 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.6 F 0.0  
5:00 220.9 F 0.0  
10:00 225.4 F 0.0  
15:00 223.6 F 0.0  
20:00 224.8 F 0.0  
25:00 225.7 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:55  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:51  
CYCLE COMPLETE

12/20/2022 10:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_