
Midmark M11 - v1.0.0
Total Cycles: 1140

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2021 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.9 F 0.0
2:00 158.8 F 2.3
4:00 183.9 F 6.0
6:00 203.0 F 10.7
8:00 214.5 F 4.9
10:00 231.8 F 10.5
12:00 231.5 F 7.3
14:00 244.1 F 13.3
16:00 243.1 F 11.9
18:00 255.7 F 18.6
20:00 267.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 28.9
2:00 272.1 F 29.0
3:00 272.2 F 28.7
4:00 272.3 F 29.2

Min 271.5 F 28.4
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 235.9 F 0.0
10:00 226.5 F 0.0
15:00 226.7 F 0.0
20:00 207.7 F 0.0
25:00 206.5 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

09/02/2021 11:40 AM
ACCEPTED: X
REJECTED:
Operator: U02