
Midmark M11 - v1.0.0
Total Cycles: 4075

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2023 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.1 F 0.0
2:00 97.2 F 0.8
4:00 109.0 F 2.9
6:00 138.0 F 5.6
8:00 181.7 F 8.6
10:00 191.0 F 12.9
12:00 205.2 F 3.9
14:00 214.9 F 7.2
16:00 225.5 F 11.1
18:00 228.7 F 6.7
20:00 239.9 F 11.7
22:00 237.3 F 9.5
24:00 249.7 F 15.3
26:00 260.3 F 21.2
28:00 270.3 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.0 F 29.1
2:00 272.2 F 28.7
3:00 271.6 F 28.7
4:00 272.6 F 29.1

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 222.0 F 0.0
10:00 225.8 F 0.0
15:00 223.6 F 0.0
20:00 222.0 F 0.0
25:00 221.7 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 28:46
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:30
CYCLE COMPLETE

12/01/2023 04:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____