
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003931
14:14 04/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:34
136.4 C 224kPa 4:34
136.4 C 226kPa 5:04
136.0 C 222kPa 5:34
136.2 C 224kPa 6:04
136.3 C 224kPa 6:34
136.3 C 222kPa 7:04
136.3 C 223kPa 7:34
136.4 C 224kPa 8:04
136.4 C 224kPa 8:04
VENTING 8:05
AIR DRYING 9:39
CYCLE COMPLETE 10:05

Digital Signature #
785456182230D376
