
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004100
8:06 26/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:11
STERILIZING 4:24
136.4 C 221kPa 4:24
136.2 C 223kPa 4:55
136.4 C 224kPa 5:25
135.8 C 217kPa 5:55
136.4 C 224kPa 6:25
136.2 C 220kPa 6:55
136.4 C 221kPa 7:25
136.4 C 223kPa 7:55
136.4 C 223kPa 7:55
VENTING 7:55
AIR DRYING 8:43
CYCLE COMPLETE 9:37

Digital Signature #
995A9EA0E29EE6AF
