
Midmark M11 - v1.0.5
Total Cycles: 725

Name: RITTER2
Sterilizer ID: V2313228
01/25/2022 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 182.5 F 2.6
4:00 190.5 F 7.4
6:00 212.9 F 13.1
8:00 225.5 F 9.0
10:00 235.3 F 9.9
12:00 242.0 F 12.8
14:00 241.6 F 11.8
16:00 254.4 F 18.4
18:00 266.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 271.9 F 29.4
2:00 272.2 F 29.9
3:00 271.5 F 29.4
4:00 271.9 F 29.6

Min 271.5 F 29.2
Max 272.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.5 F 0.0
5:00 221.7 F 0.0
10:00 236.7 F 0.0
15:00 231.9 F 0.0
20:00 215.0 F 0.0
25:00 215.8 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

01/25/2022 11:16 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____