
Midmark M11 - v1.0.5
Total Cycles: 3005

Name: RITTER2
Sterilizer ID: V2313228
10/13/2023 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 70.5 F 0.0
4:00 76.5 F 0.3
6:00 101.6 F 1.7
8:00 132.7 F 4.4
10:00 161.3 F 7.9
12:00 184.7 F 11.5
14:00 205.4 F 3.6
16:00 215.5 F 7.4
18:00 227.1 F 11.2
20:00 227.8 F 6.7
22:00 240.0 F 12.2
24:00 238.6 F 10.3
26:00 251.4 F 16.6
28:00 262.4 F 23.0
30:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.9 F 29.7
2:00 272.2 F 29.4
3:00 271.4 F 29.3
4:00 272.5 F 29.8

Min 271.2 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 199.5 F 0.0
10:00 204.9 F 0.0
15:00 203.2 F 0.0
20:00 203.4 F 0.0
25:00 201.6 F 0.0
30:00 200.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:15
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:16
CYCLE COMPLETE

10/13/2023 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____