
Midmark M11 - v1.0.5
Total Cycles: 4351

Name: RITTER2
Sterilizer ID: V2313228
02/12/2025 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 73.8 F 0.0
4:00 100.9 F 0.3
6:00 105.1 F 2.3
8:00 136.7 F 5.0
10:00 163.0 F 8.2
12:00 179.6 F 11.4
14:00 192.6 F 4.0
16:00 205.7 F 6.5
18:00 216.1 F 9.9
20:00 225.5 F 11.1
22:00 231.7 F 9.5
24:00 237.8 F 10.8
26:00 242.5 F 12.3
28:00 253.4 F 17.9
30:00 263.5 F 23.9
32:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.7 F 30.1
2:00 272.1 F 29.3
3:00 272.1 F 30.0
4:00 272.8 F 29.9

Min 271.6 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 195.0 F 0.0
10:00 203.3 F 0.0
15:00 193.3 F 0.0
20:00 182.4 F 0.0
25:00 182.9 F 0.0
30:00 193.8 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:22
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:10:50
CYCLE COMPLETE

02/12/2025 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____