
Midmark M11 - v1.0.5
Total Cycles: 3443

Name: RITTER2
Sterilizer ID: V2313228
03/07/2024 06:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 171.6 F 2.8
4:00 185.9 F 6.8
6:00 205.4 F 11.0
8:00 214.8 F 5.1
10:00 226.2 F 9.7
12:00 232.7 F 9.4
14:00 242.7 F 12.6
16:00 242.2 F 11.7
18:00 254.6 F 18.0
20:00 265.4 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.2 F 29.8
2:00 273.1 F 30.1
3:00 272.6 F 29.7
4:00 271.9 F 29.3

Min 270.9 F 28.6
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 225.2 F 0.0
10:00 230.9 F 0.0
15:00 229.8 F 0.0
20:00 230.0 F 0.0
25:00 226.3 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:01:21
CYCLE COMPLETE

03/07/2024 07:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____