
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003137
19:23 10/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:04
STERILIZING 4:11
136.4 C 224kPa 4:11
136.3 C 224kPa 4:41
136.4 C 224kPa 5:11
136.2 C 221kPa 5:41
136.2 C 223kPa 6:11
136.3 C 224kPa 6:41
136.4 C 223kPa 7:11
136.4 C 221kPa 7:41
136.4 C 221kPa 7:41
VENTING 7:41
AIR DRYING 8:34
CYCLE COMPLETE 8:51

Digital Signature #
B5DC2E513E7EB325
