
Midmark M11 - v1.0.0
Total Cycles: 2334

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2022 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.9 F 0.0
2:00 156.9 F 2.2
4:00 163.0 F 5.6
6:00 188.2 F 9.9
8:00 209.6 F 6.2
10:00 226.5 F 6.8
12:00 237.5 F 11.3
14:00 232.4 F 7.5
16:00 245.3 F 13.3
18:00 242.9 F 11.6
20:00 255.2 F 18.0
22:00 266.4 F 24.7
24:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.4 F 28.7
2:00 272.5 F 29.1
3:00 271.6 F 28.8
4:00 272.2 F 28.7

Min 271.3 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 232.7 F 0.0
10:00 235.8 F 0.0
15:00 226.5 F 0.0
20:00 217.0 F 0.0
25:00 215.4 F 0.0
30:00 234.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

09/19/2022 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____