

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3355

Name: RITTER2  
Sterilizer ID: V2313228  
02/05/2024 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.9 F 0.0  
2:00 140.5 F 1.0  
4:00 153.2 F 3.9  
6:00 178.0 F 7.9  
8:00 198.8 F 12.2  
10:00 218.0 F 6.0  
12:00 229.7 F 10.6  
14:00 230.2 F 7.7  
16:00 242.6 F 13.3  
18:00 242.9 F 12.3  
20:00 255.2 F 18.7  
22:00 266.7 F 25.6  
24:00 273.0 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.5  
1:00 273.1 F 30.1  
2:00 271.7 F 29.4  
3:00 272.7 F 30.2  
4:00 272.3 F 29.7

Min 271.4 F 29.3  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 198.9 F 0.0  
10:00 192.1 F 0.0  
15:00 191.5 F 0.0  
20:00 190.6 F 0.0  
25:00 189.2 F 0.0  
30:00 188.3 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 24:21  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:02:57  
CYCLE COMPLETE

02/05/2024 01:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_