
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005803
13:23 20/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:05
STERILIZING 4:08
136.4 C 220kPa 4:08
136.4 C 226kPa 4:38
136.4 C 224kPa 5:08
136.3 C 224kPa 5:38
136.3 C 222kPa 6:08
136.1 C 222kPa 6:38
136.4 C 225kPa 7:08
136.4 C 225kPa 7:38
136.4 C 225kPa 7:38
VENTING 7:38
AIR DRYING 8:27
CYCLE COMPLETE 8:37

Digital Signature #
E72EA7C0093A31DE
