
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000109
8:57 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 4:13
STERILIZING 5:38
136.0 C 222kPa 5:38
136.4 C 226kPa 6:08
136.3 C 225kPa 6:38
136.3 C 225kPa 7:08
136.4 C 224kPa 7:38
136.3 C 225kPa 8:08
136.4 C 225kPa 8:38
136.2 C 224kPa 9:08
136.2 C 224kPa 9:08
VENTING 9:08
AIR DRYING 10:11
CYCLE COMPLETE 25:11

Digital Signature #
9801E85EB7D714F9
