

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2516

Name: RITTER2  
Sterilizer ID: V2313228  
06/14/2023 12:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
2:00 192.0 F 4.0  
4:00 202.6 F 9.7  
6:00 228.8 F 7.2  
8:00 235.6 F 11.8  
10:00 236.4 F 9.7  
12:00 236.6 F 9.0  
14:00 249.3 F 15.5  
16:00 261.6 F 22.4  
18:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.6  
1:00 272.9 F 29.8  
2:00 272.6 F 30.2  
3:00 271.6 F 29.5  
4:00 273.1 F 30.1  
  
Min 271.5 F 29.5  
Max 273.2 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
5:00 252.9 F 0.0  
10:00 234.8 F 0.0  
15:00 225.7 F 0.0  
20:00 215.7 F 0.0  
25:00 229.2 F 0.0  
30:00 230.6 F 0.0  
  
Drying Complete

FILLING: 2:08  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

06/14/2023 01:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_