
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000952
10:36 25/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:42
STERILIZING 3:54
136.2 C 220kPa 3:54
136.4 C 233kPa 4:24
136.2 C 231kPa 4:54
136.4 C 232kPa 5:24
136.4 C 232kPa 5:54
136.2 C 230kPa 6:24
136.1 C 229kPa 6:54
136.4 C 232kPa 7:24
136.4 C 232kPa 7:24
VENTING 7:24
AIR DRYING 8:14
CYCLE COMPLETE 23:14

Digital Signature #
34418B7CE28A3220
