

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1401

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2021 01:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.3 F 0.0  
2:00 142.0 F 0.9  
4:00 164.1 F 3.4  
6:00 187.0 F 7.0  
8:00 206.8 F 11.9  
10:00 222.5 F 5.9  
12:00 234.5 F 10.7  
14:00 231.9 F 7.3  
16:00 243.5 F 12.8  
18:00 241.8 F 11.2  
20:00 254.2 F 17.7  
22:00 265.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.8 F 29.3  
2:00 272.5 F 28.8  
3:00 271.2 F 28.4  
4:00 272.4 F 29.3

Min 271.1 F 28.3  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 226.3 F 0.0  
10:00 220.6 F 0.0  
15:00 221.8 F 0.0  
20:00 215.6 F 0.0  
25:00 223.9 F 0.0  
30:00 244.8 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 23:47  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:02:26  
CYCLE COMPLETE

11/29/2021 02:55 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02