

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001319  
14:51 26/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 3:02  
STERILIZING 4:09  
136.0 C 220kPa 4:09  
136.4 C 224kPa 4:39  
136.4 C 225kPa 5:09  
136.4 C 221kPa 5:39  
136.1 C 224kPa 6:09  
136.0 C 221kPa 6:39  
135.9 C 220kPa 7:09  
136.3 C 225kPa 7:39  
136.3 C 225kPa 7:39  
VENTING 7:39  
AIR DRYING 8:31  
CYCLE COMPLETE 10:38  
-----

Digital Signature #  
8E74E989BDD211FE  
-----