
Midmark M11 - v1.0.0
Total Cycles: 4209

Name: RITTER 1
Sterilizer ID: V2178198
01/16/2024 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.6 F 0.0
2:00 87.6 F 0.5
4:00 104.4 F 2.1
6:00 135.2 F 4.7
8:00 160.7 F 7.6
10:00 183.1 F 10.8
12:00 199.1 F 4.9
14:00 216.6 F 4.8
16:00 224.3 F 7.9
18:00 231.9 F 11.8
20:00 231.8 F 7.3
22:00 241.2 F 12.4
24:00 239.4 F 10.2
26:00 251.1 F 16.0
28:00 261.5 F 21.9
30:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.1
2:00 271.9 F 28.5
3:00 272.1 F 29.0
4:00 272.6 F 29.0

Min 271.5 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 209.8 F 0.0
10:00 222.2 F 0.0
15:00 219.9 F 0.0
20:00 218.3 F 0.0
25:00 221.2 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:09:21
CYCLE COMPLETE

01/16/2024 03:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____