

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4315

Name: RITTER2  
Sterilizer ID: V2313228  
01/29/2025 07:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.5 F 0.0  
2:00 175.5 F 3.9  
4:00 201.5 F 9.7  
6:00 218.5 F 5.9  
8:00 233.4 F 11.7  
10:00 235.3 F 9.6  
12:00 235.7 F 8.6  
14:00 250.2 F 16.0  
16:00 263.1 F 23.4  
18:00 271.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.1 F 29.2  
2:00 272.7 F 29.6  
3:00 272.0 F 29.5  
4:00 272.0 F 29.1  
  
Min 271.5 F 28.9  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
5:00 229.3 F 0.0  
10:00 219.7 F 0.0  
15:00 216.2 F 0.0  
20:00 218.8 F 0.0  
25:00 220.0 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 2:17  
HEATING: 18:01  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:34  
CYCLE COMPLETE

01/29/2025 08:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_