

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1938

Name: RITTER 1  
Sterilizer ID: V2178198  
05/12/2022 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
2:00 198.9 F 3.0  
4:00 211.9 F 8.2  
6:00 219.8 F 9.7  
8:00 230.4 F 8.4  
10:00 240.5 F 13.6  
12:00 239.4 F 10.8  
14:00 237.6 F 8.7  
16:00 250.7 F 15.6  
18:00 262.4 F 22.3  
20:00 272.8 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.2  
1:00 271.5 F 28.6  
2:00 272.6 F 28.9  
3:00 271.7 F 28.3  
4:00 271.3 F 28.4  
  
Min 271.1 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 231.6 F 0.0  
10:00 234.9 F 0.0  
15:00 226.9 F 0.0  
20:00 221.9 F 0.0  
25:00 222.5 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

05/12/2022 11:07 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02