

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001853  
14:15 15/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:34  
STERILIZING 3:34  
136.0 C 221kPa 3:34  
136.3 C 226kPa 4:04  
136.3 C 226kPa 4:34  
136.4 C 227kPa 5:04  
136.1 C 223kPa 5:34  
136.3 C 226kPa 6:04  
136.3 C 226kPa 6:34  
136.4 C 227kPa 7:04  
136.4 C 227kPa 7:04  
VENTING 7:04  
AIR DRYING 7:59  
CYCLE COMPLETE 22:42  
-----

Digital Signature #  
BB1B0D3F88A4BDFC  
-----