

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002092  
11:08 24/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:17  
STERILIZING 4:28  
136.4 C 226kPa 4:28  
136.2 C 224kPa 4:58  
136.4 C 226kPa 5:28  
136.3 C 224kPa 5:58  
136.0 C 222kPa 6:28  
136.4 C 225kPa 6:58  
136.4 C 225kPa 7:28  
136.3 C 225kPa 7:58  
136.3 C 225kPa 7:58  
VENTING 7:58  
AIR DRYING 8:52  
CYCLE COMPLETE 9:45  
-----

Digital Signature #  
77870CD2CAC061B5  
-----