

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2960

Name: RITTER 1  
Sterilizer ID: V2178198  
03/01/2023 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
2:00 197.0 F 2.5  
4:00 190.7 F 6.8  
6:00 211.1 F 12.1  
8:00 224.9 F 7.1  
10:00 236.2 F 12.5  
12:00 237.3 F 10.0  
14:00 237.2 F 8.1  
16:00 250.0 F 15.3  
18:00 262.0 F 22.2  
20:00 272.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 29.1  
2:00 271.6 F 28.4  
3:00 272.4 F 29.3  
4:00 272.3 F 28.9

Min 271.4 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 215.5 F 0.0  
10:00 234.3 F 0.0  
15:00 227.1 F 0.0  
20:00 231.9 F 0.0  
25:00 232.0 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:22  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

03/01/2023 11:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_