

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3276

Name: RITTER 1  
Sterilizer ID: V2178198  
05/24/2023 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
2:00 194.3 F 1.4  
4:00 214.0 F 6.9  
6:00 224.2 F 12.5  
8:00 231.6 F 8.4  
10:00 241.1 F 11.2  
12:00 243.9 F 12.4  
14:00 242.8 F 11.5  
16:00 256.2 F 18.5  
18:00 267.8 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.7 F 28.9  
2:00 271.7 F 28.2  
3:00 272.4 F 28.9  
4:00 272.4 F 28.7  
  
Min 271.4 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 228.3 F 0.0  
10:00 245.1 F 0.0  
15:00 242.1 F 0.0  
20:00 234.6 F 0.0  
25:00 225.0 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 19:20  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:46  
CYCLE COMPLETE

05/24/2023 01:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_