
Midmark M11 - v1.0.5
Total Cycles: 3413

Name: RITTER2
Sterilizer ID: V2313228
02/26/2024 04:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.1 F 0.0
2:00 88.6 F 0.0
4:00 95.6 F 0.3
6:00 126.8 F 1.9
8:00 153.5 F 4.9
10:00 176.8 F 8.2
12:00 195.2 F 11.8
14:00 217.4 F 5.4
16:00 225.9 F 9.4
18:00 231.7 F 11.3
20:00 235.1 F 10.9
22:00 235.0 F 8.8
24:00 248.0 F 14.8
26:00 258.4 F 20.8
28:00 268.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 29.8
1:00 272.1 F 29.7
2:00 271.9 F 29.8
3:00 271.7 F 30.0
4:00 271.8 F 30.1

Min 271.6 F 29.7
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 234.1 F 0.0
10:00 225.8 F 0.0
15:00 227.7 F 0.0
20:00 214.3 F 0.0
25:00 217.5 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

02/26/2024 06:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____