

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004500  
15:19 05/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:33  
STERILIZING 3:30  
136.4 C 225kPa 3:30  
136.4 C 226kPa 4:00  
136.4 C 222kPa 4:30  
136.4 C 222kPa 5:00  
136.2 C 223kPa 5:30  
136.4 C 224kPa 6:00  
136.4 C 225kPa 6:30  
136.4 C 224kPa 7:00  
136.4 C 224kPa 7:00  
VENTING 7:00  
AIR DRYING 7:55  
CYCLE COMPLETE 9:25

-----  
Digital Signature #  
DC475B9FFE899A81  
-----