

-----  
Midmark M11 - v1.0.0  
Total Cycles: 674

Name: MIDMARK  
Sterilizer ID: V2178198  
02/17/2021 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.2 F 0.0  
2:00 183.4 F 3.2  
4:00 189.2 F 8.1  
6:00 234.5 F 9.1  
8:00 226.1 F 10.3  
10:00 232.8 F 7.5  
12:00 239.7 F 10.3  
14:00 247.0 F 13.6  
16:00 260.5 F 21.0  
18:00 271.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.6 F 29.1  
2:00 272.3 F 29.2  
3:00 272.1 F 29.0  
4:00 271.9 F 29.0  
  
Min 271.1 F 28.5  
Max 273.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
5:00 234.8 F 0.0  
10:00 225.7 F 0.0  
15:00 229.6 F 0.0  
20:00 226.9 F 0.0  
25:00 228.4 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 18:34  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 00:57:16  
CYCLE COMPLETE

02/17/2021 03:33 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01