
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006875
8:36 31/01/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:00
STERILIZING 4:05
136.4 C 223kPa 4:05
136.4 C 224kPa 4:35
136.4 C 218kPa 5:05
136.4 C 223kPa 5:35
136.1 C 221kPa 6:05
136.4 C 223kPa 6:35
136.2 C 219kPa 7:05
136.3 C 222kPa 7:35
136.3 C 222kPa 7:35
VENTING 7:35
AIR DRYING 8:24
CYCLE COMPLETE 8:34

Digital Signature #
514BD6AB06E92A35
