
Midmark M11 - v1.0.0
Total Cycles: 4077

Name: RITTER 1
Sterilizer ID: V2178198
12/04/2023 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.5 F 0.0
2:00 105.6 F 0.6
4:00 121.8 F 2.5
6:00 148.4 F 5.2
8:00 169.8 F 8.1
10:00 188.4 F 11.4
12:00 207.4 F 3.5
14:00 214.8 F 6.7
16:00 224.6 F 10.5
18:00 229.8 F 8.1
20:00 237.9 F 11.0
22:00 234.9 F 8.5
24:00 247.9 F 14.5
26:00 258.4 F 20.3
28:00 268.4 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.3
2:00 271.5 F 28.6
3:00 272.3 F 29.4
4:00 272.1 F 29.0

Min 271.3 F 28.6
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 212.1 F 0.0
10:00 216.3 F 0.0
15:00 214.1 F 0.0
20:00 210.0 F 0.0
25:00 208.8 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:17
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:07:53
CYCLE COMPLETE

12/04/2023 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____