

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3707

Name: RITTER2  
Sterilizer ID: V2313228  
06/17/2024 02:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.0 F 0.0  
2:00 100.5 F 0.2  
4:00 113.2 F 1.9  
6:00 138.6 F 4.6  
8:00 165.8 F 8.0  
10:00 182.4 F 11.8  
12:00 213.1 F 5.3  
14:00 218.6 F 9.5  
16:00 231.3 F 10.6  
18:00 235.4 F 10.6  
20:00 235.0 F 8.4  
22:00 246.3 F 14.3  
24:00 257.3 F 20.2  
26:00 267.4 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 271.3 F 29.9  
2:00 272.2 F 30.2  
3:00 272.2 F 30.4  
4:00 272.3 F 30.5

Min 270.9 F 29.3  
Max 273.1 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
5:00 210.0 F 0.0  
10:00 207.6 F 0.0  
15:00 202.9 F 0.0  
20:00 208.9 F 0.0  
25:00 205.1 F 0.0  
30:00 203.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 27:24  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:05:58  
CYCLE COMPLETE

06/17/2024 03:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_