

-----  
Midmark M11 - v1.0.5  
Total Cycles: 52

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2021 02:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.2 F 0.0  
2:00 169.0 F 3.4  
4:00 183.7 F 7.6  
6:00 204.7 F 12.3  
8:00 217.8 F 6.7  
10:00 229.2 F 11.6  
12:00 234.2 F 9.3  
14:00 240.0 F 10.4  
16:00 246.9 F 14.3  
18:00 258.6 F 20.5  
20:00 268.8 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 271.5 F 29.3  
2:00 272.0 F 29.8  
3:00 272.8 F 30.2  
4:00 272.7 F 29.8

Min 271.0 F 28.8  
Max 272.9 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 218.5 F 0.0  
10:00 227.7 F 0.0  
15:00 229.5 F 0.0  
20:00 229.2 F 0.0  
25:00 203.8 F 0.0  
30:00 197.5 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 21:23  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:00:33  
CYCLE COMPLETE

05/27/2021 03:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_