

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1199

Name: RITTER2  
Sterilizer ID: V2313228  
06/20/2022 07:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.9 F 0.0  
2:00 75.6 F 0.0  
4:00 79.4 F 0.1  
6:00 98.8 F 2.4  
8:00 133.1 F 5.5  
10:00 163.3 F 9.2  
12:00 185.4 F 12.9  
14:00 210.3 F 4.3  
16:00 220.7 F 7.9  
18:00 230.7 F 11.7  
20:00 230.5 F 7.7  
22:00 241.2 F 12.8  
24:00 240.0 F 10.9  
26:00 251.5 F 16.8  
28:00 262.2 F 22.9  
30:00 271.8 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.6  
1:00 272.3 F 29.6  
2:00 272.4 F 30.0  
3:00 272.3 F 30.2  
4:00 271.9 F 29.5

Min 271.8 F 29.5  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 199.7 F 0.0  
10:00 188.0 F 0.0  
15:00 185.8 F 0.0  
20:00 179.2 F 0.0  
25:00 176.8 F 0.0  
30:00 179.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:31  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:09:27  
CYCLE COMPLETE

06/20/2022 08:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_