
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000798
15:53 11/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:11
STERILIZING 4:24
136.0 C 220kPa 4:24
136.4 C 226kPa 4:54
136.4 C 225kPa 5:24
136.1 C 222kPa 5:54
136.4 C 224kPa 6:24
136.4 C 225kPa 6:54
136.0 C 221kPa 7:24
136.3 C 222kPa 7:54
136.3 C 222kPa 7:54
VENTING 7:54
AIR DRYING 8:50
CYCLE COMPLETE 8:57

Digital Signature #
D661727707BDBB8C
