
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000083
10:54 21/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:18
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:18

Digital Signature #
5B047F2299810DDA
