
Midmark M11 - v1.0.5
Total Cycles: 687

Name: RITTER2
Sterilizer ID: V2313228
01/12/2022 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.9 F 0.0
2:00 202.7 F 3.8
4:00 206.2 F 9.5
6:00 222.0 F 7.2
8:00 233.3 F 11.3
10:00 235.1 F 9.6
12:00 237.2 F 9.2
14:00 250.5 F 16.2
16:00 263.0 F 23.4
18:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.5 F 29.7
2:00 271.7 F 29.5
3:00 272.6 F 29.7
4:00 271.5 F 29.2

Min 271.5 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 227.2 F 0.0
10:00 230.4 F 0.0
15:00 222.3 F 0.0
20:00 216.7 F 0.0
25:00 211.1 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:29
CYCLE COMPLETE

01/12/2022 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____