

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002503  
13:30 10/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:13  
STERILIZING 4:16  
136.4 C 224kPa 4:16  
136.3 C 225kPa 4:46  
136.2 C 225kPa 5:16  
136.3 C 226kPa 5:46  
136.4 C 226kPa 6:16  
136.2 C 223kPa 6:46  
136.4 C 227kPa 7:16  
136.2 C 224kPa 7:46  
136.2 C 224kPa 7:46  
VENTING 7:47  
AIR DRYING 8:46  
CYCLE COMPLETE 15:43

-----  
Digital Signature #  
80C6A123F2967E1B  
-----