

-----  
Midmark M11 - v1.0.5  
Total Cycles: 719

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2022 07:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.7 F 0.0  
2:00 77.3 F 0.2  
4:00 107.0 F 1.1  
6:00 146.2 F 3.1  
8:00 174.5 F 5.9  
10:00 181.2 F 9.6  
12:00 189.1 F 7.2  
14:00 212.5 F 5.9  
16:00 223.3 F 9.9  
18:00 230.9 F 9.6  
20:00 236.1 F 10.9  
22:00 236.1 F 8.6  
24:00 247.9 F 15.0  
26:00 259.2 F 21.3  
28:00 269.3 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.6 F 30.1  
2:00 271.3 F 29.4  
3:00 272.6 F 30.2  
4:00 271.6 F 29.4

Min 271.2 F 29.3  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.5 F 0.0  
5:00 195.5 F 0.0  
10:00 194.0 F 0.0  
15:00 190.5 F 0.0  
20:00 187.8 F 0.0  
25:00 187.2 F 0.0  
30:00 206.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 29:04  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:07:03  
CYCLE COMPLETE

01/24/2022 08:44 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_