
Midmark M11 - v1.0.0
Total Cycles: 5078

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2024 02:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.4 F 0.0
2:00 94.3 F 0.0
4:00 108.1 F 0.0
6:00 148.0 F 0.7
8:00 175.3 F 0.7
10:00 194.4 F 1.9
12:00 209.4 F 4.9
14:00 215.9 F 8.2
16:00 226.5 F 11.9
18:00 228.6 F 6.9
20:00 238.7 F 11.8
22:00 235.6 F 9.0
24:00 241.2 F 10.6
26:00 246.0 F 13.3
28:00 257.1 F 19.2
30:00 267.3 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 29.0
2:00 271.6 F 28.4
3:00 272.7 F 29.0
4:00 271.7 F 28.3

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 204.2 F 0.0
10:00 224.3 F 0.0
15:00 232.3 F 0.0
20:00 226.2 F 0.0
25:00 223.3 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:32
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:10:10
CYCLE COMPLETE

11/11/2024 03:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____