

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3730

Name: RITTER 1  
Sterilizer ID: V2178198  
09/07/2023 07:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.0 F 0.0  
2:00 76.7 F 0.0  
4:00 80.8 F 0.0  
6:00 107.5 F 0.0  
8:00 141.7 F 1.4  
10:00 169.9 F 3.4  
12:00 195.1 F 6.1  
14:00 204.4 F 9.7  
16:00 217.0 F 13.4  
18:00 222.9 F 6.8  
20:00 233.8 F 10.9  
22:00 232.2 F 7.4  
24:00 243.4 F 12.7  
26:00 241.2 F 11.0  
28:00 253.7 F 17.2  
30:00 264.4 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.4 F 28.8  
2:00 271.7 F 28.5  
3:00 272.2 F 28.7  
4:00 271.6 F 28.5  
  
Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
5:00 204.1 F 0.0  
10:00 206.5 F 0.0  
15:00 212.6 F 0.0  
20:00 212.6 F 0.0  
25:00 220.0 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 32:00  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:10:02  
CYCLE COMPLETE

09/07/2023 09:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_