
Midmark M11 - v1.0.5
Total Cycles: 1324

Name: RITTER2
Sterilizer ID: V2313228
07/20/2022 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.3 F 0.0
2:00 139.5 F 0.0
4:00 160.9 F 0.2
6:00 176.0 F 2.7
8:00 206.6 F 6.8
10:00 221.0 F 11.3
12:00 227.2 F 6.8
14:00 236.6 F 11.9
16:00 235.6 F 9.4
18:00 237.2 F 9.8
20:00 246.5 F 14.2
22:00 258.5 F 20.6
24:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 271.2 F 29.2
2:00 272.5 F 30.3
3:00 272.7 F 30.0
4:00 272.3 F 29.5

Min 270.9 F 28.8
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 220.1 F 0.0
10:00 222.6 F 0.0
15:00 227.8 F 0.0
20:00 223.5 F 0.0
25:00 228.3 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 25:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:03:10
CYCLE COMPLETE

07/20/2022 01:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____