
Midmark M11 - v1.0.0
Total Cycles: 4113

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2023 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.1 F 0.0
2:00 91.6 F 0.6
4:00 109.8 F 2.3
6:00 147.2 F 5.0
8:00 203.5 F 8.4
10:00 219.6 F 12.5
12:00 215.6 F 3.9
14:00 223.4 F 7.4
16:00 233.0 F 11.3
18:00 227.9 F 6.7
20:00 241.4 F 12.1
22:00 237.7 F 9.9
24:00 250.7 F 15.9
26:00 261.7 F 22.2
28:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 272.5 F 29.3
2:00 271.7 F 28.8
3:00 272.5 F 29.4
4:00 271.8 F 28.7

Min 271.6 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 205.9 F 0.0
10:00 215.0 F 0.0
15:00 218.9 F 0.0
20:00 220.2 F 0.0
25:00 224.8 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 28:35
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

12/13/2023 04:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____