
Midmark M11 - v1.0.0
Total Cycles: 3247

Name: RITTER 1
Sterilizer ID: V2178198
05/17/2023 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 81.1 F 0.0
4:00 106.8 F 0.0
6:00 124.5 F 0.0
8:00 151.4 F 1.3
10:00 177.0 F 4.5
12:00 194.2 F 7.8
14:00 209.9 F 11.8
16:00 218.1 F 4.8
18:00 227.8 F 9.0
20:00 237.9 F 13.4
22:00 237.4 F 10.1
24:00 236.3 F 8.5
26:00 247.7 F 14.1
28:00 259.4 F 20.5
30:00 270.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.1 F 29.0
4:00 272.2 F 28.8

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.3 F 0.0
5:00 200.6 F 0.0
10:00 212.0 F 0.0
15:00 217.1 F 0.0
20:00 219.0 F 0.0
25:00 228.1 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:58
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:01
CYCLE COMPLETE

05/17/2023 08:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____