
Midmark M11 - v1.0.0
Total Cycles: 3031

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2023 09:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 163.9 F 3.1
4:00 168.7 F 6.3
6:00 190.2 F 10.5
8:00 214.5 F 4.9
10:00 227.3 F 9.3
12:00 233.0 F 9.4
14:00 239.7 F 11.4
16:00 237.3 F 9.3
18:00 251.0 F 15.9
20:00 262.4 F 22.7
22:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.2 F 28.9
2:00 272.7 F 29.4
3:00 271.6 F 28.6
4:00 272.5 F 29.2

Min 271.6 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 224.3 F 0.0
10:00 235.5 F 0.0
15:00 234.8 F 0.0
20:00 226.3 F 0.0
25:00 225.8 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

03/24/2023 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____