
Midmark M11 - v1.0.5
Total Cycles: 354

Name: RITTER2
Sterilizer ID: V2313228
09/13/2021 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 176.4 F 2.7
4:00 186.8 F 6.9
6:00 206.6 F 12.2
8:00 225.1 F 7.0
10:00 237.5 F 12.4
12:00 236.8 F 10.1
14:00 237.1 F 8.9
16:00 249.7 F 15.6
18:00 261.5 F 22.3
20:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.0 F 29.5
2:00 271.2 F 29.2
3:00 271.8 F 29.5
4:00 272.3 F 29.7

Min 270.9 F 28.8
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 201.7 F 0.0
10:00 198.9 F 0.0
15:00 200.1 F 0.0
20:00 198.6 F 0.0
25:00 197.0 F 0.0
30:00 206.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:29
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

09/13/2021 10:25 AM
ACCEPTED: X
REJECTED:
Operator: U01